

DILASOFT

Instructions for Use

GENERAL INFORMATION


Content

A sterile hygroscopic cervical dilator packed in a printed composite primary peel-open pouch, a piece of Instructions for use.

The DILASOFT is available in boxes of 10 dilators and in the following dimensions: 4x65mm, 4x55mm, 3x55mm.

Device description and performance

Hygroscopic cervical dilator consists of the dilating part, the polypropylene handle and the marker string (see the figure below). The dilating part is manufactured from an anisotropic hydrogel AQUACRYL. The dilator increases in diameter as it absorbs moisture from the genital tract. The marker string is tied securely to the handle of the DILASOFT to indicate its location.



1. Dilating part made of hydrogel
2. Collar
3. Marker string
4. Handle
5. Point of maximal insertion

Handling, transport, storage and waste management

Store between +15 °C and +30 °C.
Keep away from direct sunlight and high humidity.
Do not freeze.

The product, its waste materials and other consumables used during the procedure, should be disposed in accordance with local/national regulations.

Sterilization and expiration

The sterility of each device is guaranteed only when the primary packaging is unopened and undamaged.

The sterilization procedure that has been applied is marked on the label of the device – using irradiation.

INTENDED PURPOSE

Indications

The DILASOFT is for use wherever cervical softening and dilation are desired, such as cervical ripening prior to labour induction, cervical preparation prior to termination of pregnancy or other instrumentation inside the uterine cavity.

Patient target group

The DILASOFT is targeted for women indicated to labour induction or intrauterine procedure with necessary cervical ripening and/or dilation.

Intended users

The DILASOFT is for use by healthcare professionals trained in obstetrics and gynecology only.

Contraindications

The DILASOFT is contraindicated in the presence of clinically apparent genital tract infection.

WARNINGS

The DILASOFT is a single-use medical device. Instructions for its use and handling are attached to minimize exposure to conditions that may jeopardise the product, patient or user.

Re-use / re-sterilization / reprocessing¹⁾ of the DILASOFT single-use medical device may result in physical damage to the medical device, failure of intended use of the medical device, and illness or injury to the patient as a result of infection, inflammation and / or disease due to product contamination, infections and insufficient sterility of the product.

¹⁾ A process carried out on a used device in order to allow its safe reuse.

Careful placement of the device is essential to avoid traumatic injury to the cervix or uterus and to avoid migration of the device either upward into the uterus or downward into the vagina.

The DILASOFT may fragment if removed using an incorrect technique. This may result in pieces of the device being retained in the uterus. Carefully follow the Removal instructions.

Do not use if primary packaging has been opened or damaged.

Do not use if the device is not flexible (dry and hard).

Do not re-use, single use only.

Do not re-sterilize this device by any method.

Do not store at a temperature lower than +15 °C and higher than +30 °C.

Keep away from direct sunlight and high humidity.

Disposable, discard after use.

All instructions must be carefully read **prior** to using the DILASOFT.

PRECAUTIONS

As with the use of any medical device, a careful evaluation and clinical judgement should be made by the healthcare professional before using the device for the procedure to decide on the benefit/risk ratio. Alternative treatment should be considered for patients with a pre-existing condition listed under contraindications above.

Treatment options and potential risks associated with using the DILASOFT for planned procedure should be discussed with the patient before the procedure. The patient should be instructed to report any excessive bleeding, pain, or increase in temperature. The patient should be instructed to avoid bathing, vaginal douching and to refrain from sexual intercourse while the DILASOFT is in place.

The patient should be instructed that it is necessary to return for removal of the DILASOFT at the indicated time. Under no circumstances should the patient attempt to remove the DILASOFT herself.

The device should not be left in place for more than 24 hours.

If the dilator has been inserted during a procedure for termination of pregnancy, the procedure of termination of pregnancy should always be completed. In the event that the procedure of termination of pregnancy is not completed under these circumstances, the subsequent effect of dilation by DILASOFT on the fetus has not been clinically investigated.

Minor difference in flexibility and color among individual pieces may occur but the safety or effectivity of the device is not influenced.

Risks associated with the procedure

Twisting the device during its removal may cause the device to break.

In case of breakage, every attempt must be made to remove all fragments from the uterus. All fragments that are removed should be checked to ensure complete evacuation of the cavity. If in doubt, a hysteroscopy or ultrasound scan should be performed. The clinical effects of fragments retained in the genital tract are unknown.

Any cervical manipulation may cause a vaso-vagal reaction. Patients should be monitored for evidence of any unusual pallor, nausea, vertigo or weakness. By remaining recumbent for 3 to 10 minutes these symptoms usually disappear.

Complications

The following complications may be associated with use of the DILASOFT device, or may occur during the indicated procedure:

- Device entrapment
- Fragmentation or detachment of the handle
- Device expulsion
- Device retraction into the uterus
- Patient discomfort or bleeding during and/or after insertion
- Spontaneous rupture of membranes
- Spontaneous onset of labour
- Cervical laceration

USE

Examine the label of the unopened pouch and expiry date of the dilator.

Instructions for insertion

1. Insert a bivalve speculum and prepare the vagina and cervix with an antiseptic solution.
2. Remove the DILASOFT from the pouch using sterile technique.
3. Moisten the DILASOFT with sterile water or saline to lubricate the surface prior to insertion.
4. If necessary, use an appropriate technique to view the cervix and straighten the cervical canal for easier insertion of the DILASOFT.
5. Insert the DILASOFT into the cervical canal gradually and without undue force. It is important that the DILASOFT passes the internal os.
6. Do not insert the DILASOFT past the handle. The border of the collar should rest at the external os.
7. More than one DILASOFT may be inserted into the cervical canal as deemed appropriate by the healthcare professional following clinical judgement of the risk/benefit ratio.
8. When using several dilators, repeat steps 2 to 4. As many dilators as needed to achieve the desired effect should be inserted. Specific number of pieces always depends on decision and clinical judgement of healthcare professional and indications.
9. Insert a gauze pad moistened with sterile water or saline to help keep the DILASOFT in place, if needed.

Removal instructions

1. Remove vaginal packing first, if used during the insertion procedure.
2. Carefully remove the DILASOFT by grasping the handle or pulling the string. Do not twist²⁾ the DILASOFT during removal. Do not grasp the collar with forceps. Do not grasp the marker string with a sharp-edged instrument³⁾.

²⁾ Neither grasp the collar with forceps to remove the device nor twist handle when attempting to remove the device, as this may cause the device to break.

³⁾ Do not grasp the marker string with a sharp-edged instrument to remove the device, as this may cause the string to tear.

When difficulties occur during removal of the device by pulling the string, do not use excessive force on the string to remove the dilator. Use a visualization technique to identify the cause of these difficulties and remove the dilator by grasping the handle.

Occasionally, it may be necessary to use forceps to grasp the handle of the DILASOFT and exert steady traction for several minutes, while the uterus is stabilized by placing an atraumatic tenaculum through the anterior lip of the cervix.

Moisten the DILASOFT with sterile water or saline thoroughly during removal, if the dilator has stuck to the tissue, or more dilators have stuck together.

In very rare cases a "tight cervix" has been known to cause "dumbeling" of the inserted DILASOFT above and/or below the internal cervical os, making it difficult to remove. This is corrected by sliding a sequence of graduated sizes of metal dilators alongside the DILASOFT and through the internal os until sufficient dilation takes place to allow easy withdrawal.

If the DILASOFT has somehow migrated or been placed outside cervical canal, it may be located using ultrasound.

NOTE: The DILASOFT is not radiopaque.

INTERACTIONS

Within clinical investigations with the DILASOFT, a broad range of licenced medications have been administered during indicated procedures. No specific interactions between drugs / medical devices and the DILASOFT have been identified to date. Using the DILASOFT does not impose any specific limitations on standard medication administered in the context of the DILASOFT indications. Information provided to particular medications should be followed properly.

External influences

No negative interactions between the DILASOFT and external influences were observed. Desired interference include ultrasound waves that can be used for location of the inserted dilator.

TESTING OUTCOMES

Clinical

Clinical trials have not demonstrated any allergic reactions to the device. However, an allergic reaction could result from hypersensitivity to any of its components.

Clinical trials have not demonstrated any infections causally related to the DILASOFT. However, in the presence of pathogens, contamination of the device during insertion is possible.

Mechanical

The amount of dilation achieved depends on the amount of time in situ. The following table is provided as a guide.

Time in situ (hours)	Expected Dilation (in mm)		
	One DILASOFT (3x55 mm)	One DILASOFT (4x55 mm)	One DILASOFT (4x65 mm)
0	2.95 – 3.80	3.95 – 4.95	3.95 – 4.95
2	6.79 – 7.98	8.01 – 9.67	8.10 – 9.83
4	7.60 – 9.44	8.95 – 11.19	9.28 – 11.25
6	8.26 – 9.80	10.01 – 11.93	10.17 – 11.73
24	8.95 – 10.65	11.83 – 13.84	11.56 – 14.14

CONTACTS AND VIGILANCE

Please report any serious incident that has occurred in relation to the DILASOFT, to the manufacturer and your competent authority.

Please report any potential and actual product deficiencies, and product quality insufficiencies associated with the use of the DILASOFT directly to your distributor and to the manufacturer.



Manufacturer:

MEDICEM Technology s.r.o.
Karlovarska trida 20, Kamenne Zehrovice
273 01, Czech Republic
Tel.: +420 317 070 370
e-mail: technology@medicem.com
http://www.medicem.com

Liability

The manufacturer holds no liability for any side effects or resulting damages, losses

or costs that may arise as a result of the incorrect handling or use of the device.



TABLE OF USED SYMBOLS

	Keep in a dry place
	Keep away from sunlight
	Store at 15 – 30 °C
	Sterile, Sterilized using irradiation
	Do not re-use
	Degrees Celsius
	Caution, Consult accompanying documents
	Do not re-sterilize
	Do not use if package is damaged
	Consult instructions for use
	Millimeters
	Batch number
	Expiration date
	Date of manufacture
	Manufacturer
	Quantity
	Piece(s)